

## HEMOSTASIS

Five level, liquid-stable products specifically designed to challenge the IL ACL TOP system's full reportable range and eliminate manual dilutions.



### VALIDATE® D-DIMER

ORDER NO. 902il

CE

**Typical Recovery:**

191 - 3,290 ng/mL DDU

**Matrix:**

Human Plasma

**Form:**

3.0 mL, Liquid-Frozen

**Storage:**

-10 to -25 C

**Open-Vial:**

4 freeze-thaws through expiration

### VALIDATE® HEPARIN

ORDER NO. 903il

CE

**Typical Recovery:**

0.11 - 1.77 IU/mL

**Matrix:**

Human Plasma

**Form:**

3.0 mL, Liquid-Frozen

**Storage:**

-10 to -25 C

**Open-Vial:**

2 freeze-thaws through expiration

### VALIDATE® FIBRINOGEN

ORDER NO. 904il

CE

**Typical Recovery:**

55 - 794 mg/dL

**Matrix:**

Human Plasma

**Form:**

2.0 mL, Liquid-Frozen

**Storage:**

-10 to -25 C

**Open-Vial:**

Single-use, 6 hours

## EASY • FAST • EFFICIENT



### Liquid, Ready-to-Use

Increase productivity and avoid the time, error and inconvenience of reconstitution. Supplied in dropper bottles, simply bring to room temperature and dispense into the sample cup, reducing the risk of contamination.



### Open-Vial Stability

Freeze-thaw cycles allow for multiple uses - cap the bottle, return to storage and use again. For Heparin and Fibrinogen, order once and receive a second, no charge shipment, in 6 months.



### Troubleshooting

VALIDATE is ideal for use after preventative maintenance or to troubleshoot across the full reportable range.



### Linearity & Peer Groups

5 Level products with free data reduction services offers instant linearity results via MSDRx with ready access to Peer Statistics.



### Quality

Products are U.S. FDA cleared or listed and are CE marked. Our facility is registered with the U.S. FDA and our QMS is certified to ISO 13485.



### Accreditation

For use at installation, or during the laboratory's routine 6-month testing interval, rely on VALIDATE to satisfy ISO 15189, CAP, JCI and local regulatory requirements, including AMR validation.